


Bisphenol A



<u>No.</u>	<u>R.Time</u>	<u>Area</u>	<u>Peak Name</u>
1	9.947	93451	Bisphenol A

Column : Inertsil ODS-3V
 5µm
 150 × 4.6 mmI.D.

Solutes : 1. Bisphenol A

Column Cat. No. :

Eluent : CH₃CN / 60mM NaClO₄(pH2.6 adjusted by H₃PO₄) =
 40/60

Flow rate : 1.0 mL/min.

Col. Temp. : 33°C

Detection : ECD 1000mV vs Ag/AgCl

Injection Vol. : 15µL

Sample : River Water spiked at 0.1mg/L of bisphenol A