


## Analysis of Amitriptyline



Column : Inertsil ODS-3V  
150 × 4.6 mmI.D.

Column Cat. No. : 5020-01801

Eluent : Methanol/20mM Phosphate buffer(pH7.0)=65/35

Flow rate : 1.5 mL/min.

Col. Temp. : 40°C

Detection : UV 254nm

Injection Vol. : 5µL

Sample :

Solutes : 1. Uracil(0.014mg/mL)  
2. n-Butyl Paraben(0.020mg/mL)  
3. Propranolol(0.48mg/mL)  
4. Dipropyl Phthalate(0.20mg/mL)  
5. Naphthalene(0.10mg/mL)  
6. Acenaphthene(0.19mg/mL)  
7. Amitriptyline(0.09mg/mL)