


## Analysis of Pyrroloquinolinequinone [PQQ] (1)



System : GL-7400 system

Solutes : 1. PQQ (1.0 mg/L)

Column : Inertsil ODS-3 (4  $\mu$ m, 150 x 4.6 mm I.D.)

Column Cat. No. :

Eluent : A) CH<sub>3</sub>OH

B) 15 mM Tetrabutylammonium bromide (pH 3.5; H<sub>3</sub>PO<sub>4</sub>)

A/B = 50/50 ,v/v

Flow rate : 1.0 mL/min

Col. Temp. : 40 °C

Detection : FL Em 360 nm Ex 455 nm (GL-7453A FL Detector)

Injection Vol. : 5  $\mu$ L

Sample :