


Analysis of Fungicides




Conditions


System : GL7700 HPLC system
Column : InertSustain C18
 (5 μ m, 150 x 4.6 mm I.D.)
Column Cat. No. : 5020-07345
Eluent : A) CH₃CN
 B) 100 mM NaClO₄ in H₂O
 A/B = 70/30, v/v
Flow Rate : 1.0 mL/min
Col. Temp. : 40 °C
Detection : UV 220 nm (PD7752 PDA Detector)
Injection Vol. : 10 μ L
Sample : Standard

Analyte:


1. Carvacrol	4.9 mg/L
2. Benzalkonium chloride (C12)	50 mg/L
3. Benzalkonium chloride (C14)	50 mg/L



1. Carvacrol



2. Benzalkonium chloride (C12)



3. Benzalkonium chloride (C14)