


Acetylene

Specification

One GC Channel



Conditions

System : Micro GC
 Column : PoraPLOT U 10 m
 Column Temp. : 80 °C
 Carrier Gas : He 190 kPa
 Inject Time : 40 msec

Analyte:

1. Acetylene (C₂H₂) 1000 ppm (v/v)
 Balance Gas N₂